

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 27023 Accredited to ISO/IEC 17025:2017	IMQ Certification (UK) Ltd trading as IMQ UK	
	Issue No: 003 Issue date: 27 August 2024	
	Noble House Linford Wood Capital Drive Milton Keynes MK14 6QP	Contact: Cecilia Cantaluppi Tel: +44 (0) 7918005615 E-Mail: info@imquk.com Website: www.imquk.com/en
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
IMQ Spa Via Quintiliano, 43 20138 - MILANO Italy Contact: Ms. Sara Mazzei	Reaction to fire tests	A
CSI Spa Cascina Traversagna, 21 20030 - Senago (MI) Italy Contact: Ms Chiara Banfi	Reaction to fire tests	B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any suitable location	Reaction to fire tests	C



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

IMQ Certification (UK) Ltd trading as IMQ UK

Issue No: 003 **Issue date:** 27 August 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Power, control and communication cables Power, control and communication cables	Reaction to fire of power, control and communication cables	EN 50575:2014+A1:2016	A, B and C
	Heat release and smoke production measurement on cables during flame spread test. Test apparatus, procedures, results	BS EN 50399:2011+A1:2016 BS EN 50399:2022	B
	Smoke emission	EN 61034-2:2005+A2:2020	B and C
	Acid & corrosive gas emission	EN 60754-2:2014	B
	Gross calorific value (Bomb calorimeter test)	EN ISO 1716:2018	B
	Test for vertical flame spread of vertically mounted bunched wires or cables. Apparatus	BS EN IEC 60332-3-10:2018+A11:2020	A, B and C
	Test for vertical flame propagation for a single insulated wire or cable - Procedure for 1 kW pre-mixed flame	EN 60332-1-2:2004+A11:2016, BS EN 60332-1-1:2004 + A1:2015, BS EN 60332-1-3:2004 + A1:2015	A, B, and C
	Tests on gases evolved during the combustion of materials from cables	BS EN 60754-1:2014 +A1:2020	B
	Determination of the amount of halogen acid gas	BS EN 60754-1:2014 +A1:2020	B
	Determination of the degree acidity of gases for materials by measuring pH and conductivity	BS EN 60754-2:2014 +A1:2020	B



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

IMQ Certification (UK) Ltd trading as IMQ UK
Issue No: 003 Issue date: 27 August 2024

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Power, control, and communication cables (cont.)	Measurement of overall dimensions	BS EN 60811-203:2012	A, B, C
	Tests under fire conditions Power, control, and communication cables. Cables for general applications in construction works subject to reaction to fire requirements. Electric cables - Extended application of test results	BS EN 50575:2014 + A1:2016	A, B
END			



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

IMQ Certification (UK) Ltd trading as IMQ UK
Issue No: 003 **Issue date:** 27 August 2024

Testing performed by the Organisation at the locations specified

Accreditation for the purpose of UK Approved Body Activity in accordance with UKCA Requirements and UKAS Publication GEN 5

<i>Directive / Regulation</i>	Conformity Assessment procedure/ Module/article	Category of products or individual products	Essential requirements: Product specification / Properties/Standards
Construction Products Regulation 2011 (retained EU law EUR 305/2011) as amended by the Construction Products (Amendment etc.) (EU Exit) Regulations 2019 and the Construction Products (Amendment etc.) (EU Exit) Regulations 2020.	Annex V – Testing Laboratory (AVCP) System 3 - including testing away from site (Article 46)	2011/284/EU Power, control and communication cables (1/3):- Power, control and communication cables (for uses subject to regulations on reaction to fire).	EN 50575:2014+A1:2016
END			



27023

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

IMQ Certification (UK) Ltd trading as IMQ UK
Issue No: 003 Issue date: 27 August 2024

Testing performed by the Organisation at the locations specified

**Accreditation for the purpose of Notified Body Activity relating to the Northern Ireland market (CE + UKNI)
taking into account EA-2/17**

Directive / Regulation	Conformity Assessment procedure/ Module/article	Category of products or individual products	Essential requirements: Product specification / Properties/Standards
Construction Products Regulation 2011 (retained EU law EUR 305/2011) as amended by the Construction Products (Amendment etc.) (EU Exit) Regulations 2019 and the Construction Products (Amendment etc.) (EU Exit) Regulations 2020	Annex V – Testing Laboratory (AVCP) System 3 - including testing away from site (Article 46)	2011/284/EU Power, control and communication cables (1/3):- Power, control and communication cables (for uses subject to regulations on reaction to fire).	EN 50575:2014+A1:2016
END			